



Product Service

# Attestation of Conformity

No. E8A 111996 0001 Rev. 00

**Holder of Certificate:** Hangzhou Supmea Automation Co.,Ltd  
Room No 526, 5th floor, Building 4  
Singapore Hangzhou Science & Technology Park  
310018 Hangzhou  
PEOPLE'S REPUBLIC OF CHINA

**Name of Object:** Flowmeters  
(Electromagnetic Flowmeter)

**Model(s):** LDG-SUP-V2

**Description of Object:**  
Rated voltage: 22-26V DC  
Rated power: 10W max  
Input current: 4-20mA(750ohm)  
Pulse-out: 30V DC max  
Protection class: Class III  
Environments: other than residential

**Tested according to:**  
EN IEC 61326-1:2021  
EN IEC 61326-2-3:2021  
EN 61326-1:2013  
EN 61326-2-3:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 64713213018401

**Date,** 2022-01-21

( Tony Liu )

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.