



Product Service

Attestation of Conformity

No. E8A 111996 0001 Rev. 00

Holder of Certificate: Hangzhou Supmea Automation Co.,Ltd
Room No 526, 5th floor, Building 4
Singapore Hangzhou Science & Technology Park
310018 Hangzhou
PEOPLE'S REPUBLIC OF CHINA

Name of Object: Flowmeters
(Electromagnetic Flowmeter)

Model(s): LDG-SUP-V2

Description of Object:
Rated voltage: 22-26V DC
Rated power: 10W max
Input current: 4-20mA(750ohm)
Pulse-out: 30V DC max
Protection class: Class III
Environments: other than residential

Tested according to:
EN IEC 61326-1:2021
EN IEC 61326-2-3:2021
EN 61326-1:2013
EN 61326-2-3:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: www.tuvsud.com/ips-cert

Test report no.: 64713213018401

Date, 2022-01-21

(Tony Liu)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.